

CuraVance™ Bordered Silicone Foam Ag Dressing

Product Description

CuraVance™ Bordered Silicone Foam Ag Dressings are multi-layer, antibacterial foam dressings with silver, moisture vapor permeable waterproof backing, absorbent Ag foam, and soft silicone wound contact surface available with and without adhesive border. A superabsorbent core and nonwoven fluid distribution layer maximize absorption and lock in fluid under compression.

A semipermeable polyurethane film outer layer acts as a liquid and microbial barrier to prevent both strikethrough and contamination while delivering a balanced moist wound healing environment with high moisture vapor and oxygen permeability. The soft, silver-impregnated foam absorbs and retains exudate while providing broad spectrum antibacterial activity against the most common bacteria associated with wound infections for up to seven days. Dressing wear time is enhanced through the use of a superabsorbent layer and fluid transfer layer that ensure maximum use of absorbent capacity and retain fluid under compression. Patient comfort is maintained by the soft silicone adhesive wound contact layer and border, which provide pain-free atraumatic removal.

Indications for use

CuraVance™ Bordered Silicone Foam Ag

Dressings are intended for the management of moderate to heavily exuding acute or chronic wounds, including venous leg ulcers, diabetic foot ulcers, pressure injury wounds, donor sites, and partial and full thickness burns.

Precautions

CuraVance™ Bordered Silicone Foam Ag Dressings are not appropriate for use on dry or lightly exuding wounds or those covered with dry necrosis. Do not use on patients with known sensitivity to silver or allergy to the dressing components. Product is provided sterile for single use. Do not use if package is open or damaged. Do not re-sterilize.

Directions for use

1. Follow wound bed preparation protocol, including wound cleansing and debridement.
2. Select a suitable dressing pad size that overlaps the wound margin by approximately two centimeters.
3. Apply pad to the wound bed and smooth down the dressing edge and adhesive border.
4. To remove, gently release dressing edge and adhesive border and remove dressing. Dispose as medical waste.
5. Follow wound bed preparation protocol, including cleansing with appropriate solution before applying subsequent dressings.

Frequency of change

Dressing change frequency will vary as wound

healing progresses. Daily dressing changes may be required initially until exudation decreases. Change when dressing is saturated, soiled or according to local protocol (up to seven days).

Warnings

Discontinue use and notify appropriate medical personnel if signs of infection are observed. Do not use with oxidizing solutions, such as hypochlorite or hydrogen peroxide (H₂O₂).

Storage

Store in a well-ventilated, cool, dry space.

Not made with natural rubber latex.



MedWay Inc.

1650 Horizon Parkway, Suite 450
Buford, GA 30518 U.S.A.

Email: sales@medwayus.com
www.medwayus.com

Made in China
™ Trademark of MedWay Group.

Revision: cv.ifu.3.1-v1